



भारतीय प्रौद्योगिकी संस्थान मुंबई

पवई, मुंबई-400 076, भारत

Indian Institute of Technology Bombay

Powai, Mumbai-400 076, India

दूरभाष/Phone : (+91-22) 2572 2545

फैक्स/Fax : (+91-22) 2572 3480

वेबसाइट/Website : www.iitb.ac.in

IIIT Bombay

10th February, 2022

Corrigendum –IV

Reference (PR No. 1000020190) RFx No. 6100000959

Dynamic Mechanical Analyzer(DMA)

Please note that, following points were modified:

Sr No	Details	Previous clause	Modified clause
1.	Bid submission End Date/Date & Time of submission (Online RFx clause)	15.02.2022 at 13.00	22.02.2022 at 13.00
2.	Bid Opening Date & Time (Online RFx clause)	15.02.2022 at 16.00	22.02.2022 at 16.00
3.	Clause No[1.25] (Technical specifications)	Sample Holders for Deformation Modes: a) Single & dual cantilever, b) 3 and 4 point bending c) Tension d) Shearing e) Compression	Sample Holders for Deformation Modes: a) Single & dual cantilever, b) 3 point bending c) Tension d) Shearing e) Compression
4.	Clause No[2.1] (Technical specifications)	Standard samples for Calibration should be provided, e.g., annealed and stress free PMMA tension, compression and bending samples.	Standard samples for verification should be supplied along with the instrument, for tension, compression, bending and shear experiments
5.	Clause No[2.6] (Technical specifications)	UV equipment for in situ curing and testing, along with related safety equipment.	<u>The clause is deleted</u>
6.	Clause No[2.7] (Technical specifications)	Dielectric Analyzer (DEA) for in situ testing	<u>The clause is deleted</u>
7.	Clause No[2.8] (Technical specifications)	Sample imaging: Capability to image the sample during experiment through an optical window under all	<u>The clause is deleted</u>

		loading conditions, with integrated software for image capture at ≥ 30 fps and > 2 Mega pixel resolution.	
8.	Clause No[5.2] (Technical specifications)	The vendor should have installed the quoted machine in India and a documentary evidence of the same and evidence of satisfactory working & after sale support to be submitted along with tender document.	The vendor should have installed the quoted machine in India/Reputed Academic and Research labs globally and should provide documentary evidence of the same; evidence of satisfactory working & after sale support has to be submitted along with the tender document; and the vendor should have Service Centre in India for providing service support in a timely manner.

Assistant Registrar
Materials Management Division